Copying Channels on Milk Analysers - Delta Lactoscope Filter

Print Off The Parameters Of The Channel To Be Copied

Select Product

↓

Press Setup

↓

Select setup from the next menu screen

↓

Select Print Products

↓

Set Dip Switch 3 inside the analyser (QCL will advise how to do this)

Creating New Channel

Select Channel to be created

↓

Select Setup

↓

Adjust Product

↓

Input password

↓

Select Next

↓

Yes to Change Correction Factors

Enter details from printout

Change Product name using attached codes

Set Pump settings the same as printout

Print to confirm

Check against original

Input Slope & Bias

Select Product

Select Product Setup

Select Adjust Product

Select Manual

Input Slope and Bias settings from printout