ISO 13485:2016

ADVANCED INSTRUMENTS

Certificate of Analysis

Product Name: Fluorophos® Alkaline Phosphatase (ALP) Reagent Set
Product Number: FLA224/FLA225
Kit Lot Number: C301189008
Buffer Lot Number: 21477
Substrate Lot Number: 21477A
Expiration Date: 2028-02
Quality Release Date: 18 March 2026

Internal Test Specification and Result

Test	Specification	Result
Alkaline Phosphatase (ALP Reagent Set – A/D Reading	< 700 FLU	649 FLU
Alkaline Phosphatase (ALP Reagent Set – Fluorometer Measurement	< 10 mU/L	< 10 mU/L

Storage/Handling/Stability

Storage	Handling	Stability
2 - 8°C (36 - 46°F)	Tightly reconstituted substrate bottle after use and immediately store at 2 - 8°C (36 - 46°F). Do not freeze.	Unopened: Refer to expiry date on the box (2 - 8°C (36 - 46°F)). Opened: Stable for 60 days when stored at 2 - 8°C (36 - 46°F)

The information on this Certificate is only valid when the manufacturer's instructions are followed.

Approval

Tested By: Mark Hunter _____ Approved By: MEREDITH PESTA

Mark Hunter 3/19/2026
(Signature) (Date)

Meredith Pesta 3/19/2026
(Signature) (Date)

Two Technology Way · Norwood, Massacusetts 02062, USA · www.aicompanies.com
781-320-9000 · 800-255-4034 · Fax 781-320-8181