ISO 13485:2016

ADVANCED INSTRUMENTS

# Certificate of Analysis

**Product Name:** Fluorophos® Calibrator Set
**Product Number:** FLA250
**Kit Lot Number:** 21476
**Calibrator A Lot Number:** 21476
**Calibrator B Lot Number:** 21476A
**Calibrator C Lot Number:** 21476B
**Expiration Date:** 2027-08
**Quality Release Date:** 30 March 2026

## Internal Test Specification and Result

| Test | Specification | Result |
|---|---|---|
| Calibration Ratio | 150.0 ± 6.0 | 151.2 |

## Storage/Handling/Stability

| Storage | Handling | Stability |
|---|---|---|
| 2 - 8°C (36 - 46°F) | Tightly close bottles after use and immediately store at 2 - 8°C (36 - 46°F).<br>Do not freeze. | Unopened: Refer to expiry date on the box (2 - 8°C (36 - 46°F)).<br>Opened: Stable for 60 days when stored at 2 - 8°C (36 - 46°F) |

The information on this Certificate is only valid when the manufacturer's instructions are followed.

## Approval

Tested By: Mark Hunter _____ Approved By: MEREDITH PESTA

Mark Hunter 3/30/2026
(Signature) (Date)

Meredith Pesta 3/30/2026
(Signature) (Date)