ISO 13485:2016

ADVANCED INSTRUMENTS

Certificate of Analysis

Product Name: **Fluorophos® Calibrator Set**
Product Number: **FLA250**
Kit Lot Number: **21878**
Calibrator A Lot Number: **21878**
Calibrator B Lot Number: **21878A**
Calibrator C Lot Number: **21878B**
Expiration Date: **2028-01**
Quality Release Date: 24 August 2026

Internal Test Specification and Result

Test	Specification	Result
Calibration Ratio	150.0 ± 6.0	148.0

Storage/Handling/Stability

Storage	Handling	Stability
2 - 8°C (36 - 46°F)	Tightly close bottles after use and immediately store at 2 - 8°C (36 - 46°F). Do not freeze.	Unopened: Refer to expiry date on the box (2 - 8°C (36 - 46°F)). Opened: Stable for 60 days when stored at 2 - 8°C (36 - 46°F)

The information on this Certificate is only valid when the manufacturer's instructions are followed.

Approval

Tested By: Carla Rosa Approved By: MEREDITH PESTA

Carla Rosa 8/24/26 (Signature) (Date)

Meredith Pesta 8/24/2026 (Signature) (Date)