ISO 13485:2016

ADVANCED INSTRUMENTS

Certificate of Analysis

Product Name:	**Fluorophos® PhosphaCheck® Pasteurization Controls**
Product Number:	**FLA260**
Kit Lot Number:	**21481B**
Positive Control Lot Number:	**21481B**
Negative Control Lot Number:	**21481A**
Normal Control Lot Number:	**21482**
Expiration Date:	**2027-08**
Quality Release Date:	18 March 2026

Internal Test Specification and Result

Test	Specification	Result
Positive Control Fluorometer Measurement	500 ± 50 mU/L	473 mU/L
Negative Control Fluorometer Measurement	< 10 mU/L	< 10 mU/L
Normal Control Fluorometer Measurement	< 30 mU/L	< 30 mU/L

Storage/Handling/Stability

Storage	Handling	Stability
2 - 8°C (36 - 46°F)	Tightly seal the vials after use to prevent evaporation. Store vials upright to prevent spills and leakage. Do not freeze.	Unopened: Refer to expiry date on the box (2 - 8°C (36 - 46°F)). Opened: Stable for 72 hours when stored at 2 - 8°C (36 - 46°F)

The information on this Certificate is only valid when the manufacturer's instructions are followed.

Approval

Tested By: Mark Hintz _____ Approved By: MEREDITH PESTA

M.D.H. 3/18/2026 _____ Meredith Pesta 3/19/2026

(Signature) (Date) (Signature) (Date)