ISO 13485:2016

ADVANCED INSTRUMENTS

Certificate of Analysis

Product Name: **Fluorophos® Daily Instrument Control**
Product Number: **FLA280**
Lot Number: **21771A**
Expiration Date: **2027-11**
Quality Release Date: 8 June 2026

Internal Test Specification and Result

Test	Specification	Result
Daily Instrument Control – A/D Reading	602 ± 6 FLU	605 FLU

Storage/Handling/Stability

Storage	Handling	Stability
2 - 8°C (36 - 46°F)	Tightly close the bottles after use and immediately store at 2 - 8°C (36 - 46°F). Do not freeze.	Unopened: Refer to expiry date on the box (2 - 8°C (36 - 46°F)). Opened: Stable for 60 days when stored at 2 - 8°C (36 - 46°F)

The information on this Certificate is only valid when the manufacturer's instructions are followed.

Approval

Tested By: Carla Rosa Approved By: Mark Hunter

Carla Rosa 06-09-2026 (Signature) (Date)

(Signature) (Date)